

इंटरनेट

मानक

Disclosure to Promote the Right To Information

Whereas the Parliament of India has set out to provide a practical regime of right to information for citizens to secure access to information under the control of public authorities, in order to promote transparency and accountability in the working of every public authority, and whereas the attached publication of the Bureau of Indian Standards is of particular interest to the public, particularly disadvantaged communities and those engaged in the pursuit of education and knowledge, the attached public safety standard is made available to promote the timely dissemination of this information in an accurate manner to the public.

“जानने का अधिकार, जीने का अधिकार”

Mazdoor Kisan Shakti Sangathan

“The Right to Information, The Right to Live”

“पुराने को छोड़ नये के तरफ”

Jawaharlal Nehru

“Step Out From the Old to the New”

IS 8047 (1976): Forceps, Extraction, Dental, Upper Molar, Children [MHD 8: Dentistry]



“ज्ञान से एक नये भारत का निर्माण”

Satyanarayan Gangaram Pitroda

“Invent a New India Using Knowledge”



“ज्ञान एक ऐसा खजाना है जो कभी चुराया नहीं जा सकता है”

Bhartrhari—Nitiśatakam

“Knowledge is such a treasure which cannot be stolen”

BLANK PAGE





Indian Standard

SPECIFICATION FOR FORCEPS, EXTRACTION, DENTAL, UPPER MOLAR, CHILDREN

1. Scope — Covers dimensional and other requirements for dental extraction forceps for use in dental surgery of children's upper molar.

2. Material

2.1 Arms — Shall be of stainless steel conforming to Designation 30Cr13 of IS: 6603-1972 'Specification for stainless steel bars and flats'.

2.2 Screws — Shall be of stainless steel conforming to Designation 20Cr13 of IS: 6603-1972.

3. Shape and Dimensions — As shown in Fig. 1.




FIG. 1 FORCEPS, EXTRACTION, DENTAL, UPPER MOLAR, CHILDREN

3.1 A deviation of ± 2.5 percent shall be allowed on all dimensions.

4. Mass — Shall be 100 ± 5 g.

5. Other Requirements — The instruments shall satisfy the requirements given in 4 to 8 of IS: 6868-1972 'General requirements for forceps, extraction, dental'.

5.1 Performance Load — Shall be 35 ± 5 kg.

Adopted 10 May 1976

© July 1976, ISI

Gr 1